



# HVACR PORTABLES

## Quick Start Guide

EN

FR


ES

PT

DE

CN

## English

 These portable instruments are used to measure the following parameters: temperature, relative humidity, pressure, airflow and air velocity.

Hereby, Sauermann Industrie SAS declares that the radio equipment type HVACR Portable is in compliance with Directive 2014/53/EU. The full text of the EU declaration of conformity is available at the following internet address: [www.sauermanngroup.com](http://www.sauermanngroup.com)

### 1 Key operation for Si-HH3, Si-PM3, Si-TT3, Si-VH3 and Si-VV3

Key operation is differs from one device to another.

#### Si-HH3

KEY	SHORT PRESS	LONG PRESS (3 s)
ON/OFF ①	Backlight ON/OFF	Device ON/OFF
HOLD ②	Hold / Min / Max / Live value	DewPoint / WetBulb / %RH & Temp
ON/OFF and HOLD ①+②	Wireless communication ON/OFF	°C / °F

#### Si-PM3

KEY	SHORT PRESS	LONG PRESS (3 s)
ON/OFF ①	Backlight ON/OFF	Device ON/OFF
HOLD ②	Hold / Min / Max / Live value	Autozero
ON/OFF and HOLD ①+②	Wireless communication ON/OFF	Pa / hPa / mbar / inH <sub>2</sub> O / mmH <sub>2</sub> O / inHg / psi / m/s / fpm

#### Si-TT3

KEY	SHORT PRESS	LONG PRESS (3 s)
ON/OFF ①	Backlight ON/OFF	Device ON/OFF
HOLD ②	Hold / Min / Max / Live value	Delta Temp / T1&T2
ON/OFF and HOLD ①+②	Wireless communication ON/OFF	°C / °F

#### Si-VH3 / Si-VV3

KEY	SHORT PRESS	LONG PRESS (3 s)
ON/OFF ①	Backlight ON/OFF	Device ON/OFF
HOLD ②	Hold / Min / Max / Live value	m/s / km/h / fpm / mph
ON/OFF and HOLD ①+②	Wireless communication ON/OFF	°C / °F

#### Power Down menu

The instrument is supplied with auto-off activated. Enter in the Power Down Menu when starting the instrument to deactivate it.

KEY	SHORT PRESS	LONG PRESS (3 s)
ON/OFF ①	-	Enter/Exit of Power Down menu
HOLD ②	Auto-off activation/deactivation	-
ON/OFF and HOLD ①+②	-	-



1 - ON/OFF key  
2 - HOLD key



## 2 Key operation for Si-T13

KEY	MAIN FUNCTION	OTHER FUNCTION
°C/°F ①	Switches between °C and °F	Sets Emissivity, High alarm and Low alarm
Mode ②	Displays successively Current emissivity / Emissivity to change / Max temperature / Min temperature / Difference between max and min temperatures / Average temperature / High alarm / Low alarm	-
Lock ③	Continuously displays measured temperatures	Sets Emissivity, High alarm and Low alarm
Trigger ④	Measures the surface temperature	Validates an Emissivity, a High alarm and a Low alarm.



To activate/deactivate the laser dot, press ① and ④ keys simultaneously.  
To get information on other functions, please read the full technical manual downloadable from the QR code at the end of this document.

## 3 Insert the batteries in the instrument

- Unscrew (all instruments except Si-T13) and remove the batteries cover ① .
- Insert 3 (all instruments except Si-T13) or 2 (only Si-T13) alkaline batteries LR03 AAA 1.5 V respecting the polarity.
- Replace and screw (all instruments except Si-T13) the batteries cover.



## 4 Download Si-HVACR Measurement MobileApp



- Flash the QR code to download **Si-HVACR Measurement MobileApp** application on your smartphone.

OR

- Write **"Si-HVACR Measurement MobileApp"** in the search bar then validate.
- Download **Si-HVACR Measurement MobileApp** application by pressing its icon.
- Follow your smartphone indications.



### 5a Connect the smartphone and the instrument (all instruments except Si-T13)

- Turn on the instrument with a long press on ON/OFF key.
- Press ON/OFF and HOLD keys to activate the wireless communication. The wireless communication icon flashes on the instrument screen.
- Launch **Si-HVACR Measurement MobileApp** application by pressing its icon.
- Press **"Search for devices"**.
- Press the line of the required device then **"Connect"**. The wireless communication icon becomes fix when the instrument is paired to the tablet or the smartphone.

## 5b Connect the smartphone and the instrument (Si-TI3 only)

The Si-TI3 is automatically turned on with wireless communication activated when inserting batteries.

- Launch **Si-HVACR Measurement MobileApp** application by pressing its icon.
- Press **“Search for devices”**.
- Press the line of the required device then **“Connect”**.



The Si-TI3 will automatically shut-off after 15 s of non-use.

## 6 Display measurement

Once the instrument is connected to the smartphone, measurements are displayed.



## 7 Emissivity (Si-TI3 only)

Emissivity is a term used to describe the energy-emitting characteristics of materials. See table below for values of emissivity of specific materials:

Aluminium	0.30	Ice	0.98
Asbestos	0.95	Iron	0.70
Asphalt	0.95	Lead	0.50
Basalt	0.70	Limestone	0.98
Brass	0.50	Oil	0.94
Brick	0.90	Paint	0.93
Carbon	0.85	Paper	0.95
Ceramic	0.95	Plastic	0.95
Concrete	0.95	Rubber	0.95
Copper	0.95	Sand	0.90
Dirt	0.94	Skin	0.98
Frozen food	0.90	Snow	0.90
Hot food	0.93	Steel	0.80
Glass	0.85	Textile	0.94
Water	0.93	Wood	0.94
Fresh foodstuffs between 0 and 5 °C			0.90

### Set emissivity

- With the device turned on.
- Press Mode key then use Lock key to increment or Unit key to decrease the emissivity value.
- The emissivity can be changed from 0.10 (10E) to 1 (100E).
- Press Mode key to confirm.

## 8 Operating temperature, protection of the instruments and information about storage and batteries

- Si-HH3 / Si-TI3 / Si-TT3: from -10 to +60°C (14 to 140°F)
- Si-PM3 / Si-VH3: from 0 to 50°C (32 to 122°F)
- Si-VV3: from 0 to 40°C (32 to 104°F)
- Si-HH3: IP20
- Si-PM3: IP40
- Si-TT3 / Si-VH3 / Si-VV3: IP54

If the instrument is stored outside the operating temperature (for example in a van, a warehouse,...) please wait for 10 minutes in its operating temperature before starting and using it.



To ensure safety, replace batteries with the same type. Not doing so runs the risk of explosion. Dispose of used batteries according to instructions.



Download the full manual  
Télécharger le manuel complet  
Descargue el manual de usuario  
Scarica il manuale completo  
Vollständiges Handbuch herunterladen  
Descargar el manual completo  
下载使用说明书全文



**Sauermann Industrie**

ZA Bernard Moulinet  
24700 Montpon  
France  
T. +33 (0)5 53 80 85 00  
services@sauermanngroup.com

**Sauermann NA**

415 Oser Avenue, Ste. P  
Hauppauge, New York 11788  
T. (+1) 631-234-7600  
F. (+1) 631-234-7605  
info.usa@sauermanngroup.com

**Sauermann GmbH**

Leibnizstraße 6  
D – 74211 Leingarten  
T. +49 (0)7131/399990  
F. +49 (0)7131/399992  
info.germany@sauermanngroup.com

**Sauermann Ibérica**

C/Albert Einstein 33.  
Planta 3. P. I. Santa Margarida II-  
08223 Terrassa (Spain)  
T. +34 931 016 975  
info.spain@sauermanngroup.com

**Sauermann UK**

Units 7-9, Trident Business Park  
Amy Johnson Way  
Blackpool - FY4 2RP  
T. +44 (0) 870 950 6378  
F. +44 (0) 870 950 6379  
info.uk@sauermanngroup.com

**Sauermann Italia srl**

Via Golini 61/10  
40024 Castel S.Pietro Terme (BO)  
T. (+39)-051-6951033  
F. (+39)-051-942254  
info.italy@sauermanngroup.com

**Sauermann Australia**

Unit 4/14 Rodborough Road, Frenchs Forest,  
NSW 2086  
T. (+612) 8880 4631

**Sauermann China**

Room 401-6, Building 8, No. 3000 Longdong Ave. Pudong  
New District  
Shanghai 201203/R.P. China  
T. +86-21-61001877  
F. +86-21-61001870  
info.china@sauermanngroup.com

**OneTemp**<sup>®</sup>  
measure | control | record  
1300 768 887  
www.onetemp.com.au